

Fusty Odors

Technique P&T
Column InertCap 1
30m×0.25mm I.D. df=0.4µm
Column Temp. 40°C (1min Hold) - 10°C/min - 220°C
Carrier Gas He
0.8mL/min
Detector MS
Sample Conc. 1ng/mL

