

32 Halocarbons

Technique GC
 Injection Split
 Column AQUATIC
 60m×0.25mm I.D. df=1.0µm
 Column Temp. 60°C - 4°C/min - 150°C
 Carrier Gas He
 200kPa
 Split 1 : 80
 Detector FID
 Injection Size 0.1µL

- | | |
|-------------------------|-------------------------------|
| 1 1,1-Dichloroethene | 17 Dibromomethane |
| 2 Methylenechloride | 18 Toluene |
| 3 t-1,2-Dichloroethene | 19 1,1,2-Trichloroethane |
| 4 1,1-Dichloroethane | 20 1,3-Dichloropropane |
| 5 2,2-Dichloropropane | 21 Tetrachloroethene |
| 6 c-1,2-Dichloroethene | 22 Dibromochloromethane |
| 7 Chloroform | 23 1,2-Dibromomethane |
| 8 Bromochloromethane | 24 Chlorobenzene |
| 9 1,1,1-Trichloroethane | 25 m-Xylene + p-Xylene |
| 10 1,1-Dichloropropene | 26 o-Xylene |
| 11 Carbontetrachloride | 27 Bromoform |
| 12 1,2-Dichloroethane | 28 1,1,2,2-Tetrachloropropane |
| 13 Benzene | 29 1,2,3-Trichloropropane |
| 14 Trichloroethene | 30 Bromobenzene |
| 15 1,2-Dichloropropane | 31 m-Chlorotoluene |
| 16 Bromodichloromethane | 32 p-Chlorotoluene |

